

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032025\
 Data File : VX045349.D
 Acq On : 20 Mar 2025 10:29
 Operator : JC/MD
 Sample : VX0320WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0320WBL01

Quant Time: Mar 21 01:42:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	67796	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	133500	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	121748	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	47437	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61741	57.253	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.500%
35) Dibromofluoromethane	5.379	113	47374	53.069	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.140%
50) Toluene-d8	8.647	98	165758	51.219	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.440%
62) 4-Bromofluorobenzene	11.079	95	55720	51.958	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032025\
Data File : VX045349.D
Acq On : 20 Mar 2025 10:29
Operator : JC/MD
Sample : VX0320WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0320WBL01

Quant Time: Mar 21 01:42:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

