

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032025\
 Data File : VX045352.D
 Acq On : 20 Mar 2025 11:44
 Operator : JC/MD
 Sample : Q1559-09DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP03-20250310DL

Quant Time: Mar 21 01:44:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	67584	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	133155	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118349	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	45287	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61600	57.301	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.600%
35) Dibromofluoromethane	5.379	113	46858	52.627	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.260%
50) Toluene-d8	8.647	98	164552	50.978	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.960%
62) 4-Bromofluorobenzene	11.079	95	54195	50.667	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.340%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.319	101	2383	3.034	ug/l	95
12) 1,1-Dichloroethene	2.306	96	1159	1.396	ug/l	97
44) Trichloroethene	7.123	130	45737	50.517	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032025\
Data File : VX045352.D
Acq On : 20 Mar 2025 11:44
Operator : JC/MD
Sample : Q1559-09DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP03-20250310DL

Quant Time: Mar 21 01:44:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

