

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032025\
 Data File : VX045353.D
 Acq On : 20 Mar 2025 12:07
 Operator : JC/MD
 Sample : Q1559-22DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D2-20250312DL

Quant Time: Mar 21 01:45:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	72808	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	141715	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127752	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	48615	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	63476	54.810	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.620%	
35) Dibromofluoromethane	5.379	113	49703	52.451	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.900%	
50) Toluene-d8	8.647	98	177305	51.611	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.220%	
62) 4-Bromofluorobenzene	11.079	95	58857	51.702	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 103.400%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.319	101	10962	12.954	ug/l	99
44) Trichloroethene	7.123	130	33503	34.769	ug/l	98
64) Tetrachloroethene	9.275	164	765	0.935	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032025\
Data File : VX045353.D
Acq On : 20 Mar 2025 12:07
Operator : JC/MD
Sample : Q1559-22DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D2-20250312DL

Quant Time: Mar 21 01:45:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

