

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032025\
 Data File : VX045354.D
 Acq On : 20 Mar 2025 12:30
 Operator : JC/MD
 Sample : Q1559-21DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D3-20250312DL

Quant Time: Mar 21 01:45:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	71455	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	139830	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	126173	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	53223	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64357	56.623	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.240%
35) Dibromofluoromethane	5.385	113	49428	52.864	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.720%
50) Toluene-d8	8.647	98	172491	50.886	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.780%
62) 4-Bromofluorobenzene	11.079	95	59332	52.822	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	105.640%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.319	101	936	1.127	ug/l	89
12) 1,1-Dichloroethene	2.306	96	386	0.440	ug/l #	77
44) Trichloroethene	7.123	130	36938	38.851	ug/l	96
64) Tetrachloroethene	9.275	164	285	0.353	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032025\
Data File : VX045354.D
Acq On : 20 Mar 2025 12:30
Operator : JC/MD
Sample : Q1559-21DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D3-20250312DL

Quant Time: Mar 21 01:45:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

