

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032025\
 Data File : VX045356.D
 Acq On : 20 Mar 2025 13:17
 Operator : JC/MD
 Sample : Q1559-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP02-20250310

Quant Time: Mar 21 05:35:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	67663	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	132656	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	121785	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	51206	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60070	55.813	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	111.620%	
35) Dibromofluoromethane	5.379	113	47034	53.024	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	106.040%	
50) Toluene-d8	8.646	98	165862	51.577	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	103.160%	
62) 4-Bromofluorobenzene	11.079	95	58691	55.077	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	110.160%	
Target Compounds						
						Qvalue
16) Acetone	2.385	43	857	1.716	ug/l #	81
44) Trichloroethene	7.134	130	2970	3.293	ug/l	64
64) Tetrachloroethene	9.274	164	390	0.500	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032025\
Data File : VX045356.D
Acq On : 20 Mar 2025 13:17
Operator : JC/MD
Sample : Q1559-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP02-20250310

Quant Time: Mar 21 05:35:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

