

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032025\
 Data File : VX045361.D
 Acq On : 20 Mar 2025 15:13
 Operator : JC/MD
 Sample : Q1582-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB02-20250313

Quant Time: Mar 21 01:49:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	71202	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	138215	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	126962	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49435	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63283	55.875	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.760%
35) Dibromofluoromethane	5.385	113	47926	51.856	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.720%
50) Toluene-d8	8.646	98	173113	51.667	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.340%
62) 4-Bromofluorobenzene	11.079	95	57728	51.994	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.980%
Target Compounds						
16) Acetone	2.379	43	2813	5.353	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032025\
Data File : VX045361.D
Acq On : 20 Mar 2025 15:13
Operator : JC/MD
Sample : Q1582-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB02-20250313

Quant Time: Mar 21 01:49:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

