

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032025\
 Data File : VX045365.D
 Acq On : 20 Mar 2025 16:46
 Operator : JC/MD
 Sample : Q1582-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT180D-20250313

Quant Time: Mar 21 03:57:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	66194	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	132635	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	121314	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51107	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60912	57.851	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.700%
35) Dibromofluoromethane	5.379	113	46162	52.049	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.100%
50) Toluene-d8	8.647	98	164463	51.150	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.300%
62) 4-Bromofluorobenzene	11.079	95	57646	54.105	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	108.200%
Target Compounds						
16) Acetone	2.392	43	761	1.558	ug/l #	89
44) Trichloroethene	7.129	130	3312	3.672	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032025\
Data File : VX045365.D
Acq On : 20 Mar 2025 16:46
Operator : JC/MD
Sample : Q1582-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT180D-20250313

Quant Time: Mar 21 03:57:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

