

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032025\
 Data File : VX045368.D
 Acq On : 20 Mar 2025 17:55
 Operator : JC/MD
 Sample : Q1582-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D3-20250313

Quant Time: Mar 21 01:52:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	62724	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	124749	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	117366	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48478	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	57798	57.930	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.860%
35) Dibromofluoromethane	5.379	113	44178	52.961	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.920%
50) Toluene-d8	8.647	98	158129	52.289	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.580%
62) 4-Bromofluorobenzene	11.079	95	55300	55.184	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	110.360%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.319	101	525636	721.028	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	1287	1.369	ug/l	91
44) Trichloroethene	7.123	130	49305	58.127	ug/l	97
45) 1,2-Dichloropropane	7.434	63	2064	2.222	ug/l	93
64) Tetrachloroethene	9.269	164	48799	64.920	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032025\
Data File : VX045368.D
Acq On : 20 Mar 2025 17:55
Operator : JC/MD
Sample : Q1582-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D3-20250313

Quant Time: Mar 21 01:52:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

