

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032025\
 Data File : VX045369.D
 Acq On : 20 Mar 2025 18:19
 Operator : JC/MD
 Sample : Q1582-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT119S1-20250313

Quant Time: Mar 21 01:52:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	66589	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	130512	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	122323	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	47028	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	60689	57.297	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.600%
35) Dibromofluoromethane	5.379	113	46381	53.147	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.300%
50) Toluene-d8	8.647	98	162619	51.399	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.800%
62) 4-Bromofluorobenzene	11.079	95	56246	53.649	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	107.300%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.319	101	1246	1.610	ug/l	93
16) Acetone	2.380	43	1104	2.246	ug/l #	86
44) Trichloroethene	7.129	130	2255	2.541	ug/l	73
64) Tetrachloroethene	9.281	164	625	0.798	ug/l #	62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032025\
Data File : VX045369.D
Acq On : 20 Mar 2025 18:19
Operator : JC/MD
Sample : Q1582-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT119S1-20250313

Quant Time: Mar 21 01:52:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

