

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032025\
 Data File : VX045371.D
 Acq On : 20 Mar 2025 19:05
 Operator : JC/MD
 Sample : Q1582-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPOW1-7-20250314

Quant Time: Mar 21 01:53:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	81071	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	156759	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	142821	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	57498	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	73166	56.737	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 113.480%	
35) Dibromofluoromethane	5.385	113	56811	54.198	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.400%	
50) Toluene-d8	8.647	98	195439	51.430	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.860%	
62) 4-Bromofluorobenzene	11.079	95	67261	53.414	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 106.820%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	1284	2.146	ug/l	91
44) Trichloroethene	7.129	130	1180	1.107	ug/l	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032025\
Data File : VX045371.D
Acq On : 20 Mar 2025 19:05
Operator : JC/MD
Sample : Q1582-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPOW1-7-20250314

Quant Time: Mar 21 01:53:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

