

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032119\
 Data File : VX008287.D
 Acq On : 21 Mar 2019 13:43
 Operator : JC/SP
 Sample : K1963-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TB05-20190314

Quant Time: Mar 22 09:01:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 13:15:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	305173	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	497880	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	423436	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	172974	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	212813	46.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.60%	
35) Dibromofluoromethane	5.50	113	161387	48.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.84%	
50) Toluene-d8	8.71	98	612679	48.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.96%	
62) 4-Bromofluorobenzene	11.13	95	196877	45.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.50%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032119\
Data File : VX008287.D
Acq On : 21 Mar 2019 13:43
Operator : JC/SP
Sample : K1963-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TB05-20190314

Quant Time: Mar 22 09:01:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 13:15:13 2019
Response via : Initial Calibration

