

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032119\
 Data File : VX008296.D
 Acq On : 21 Mar 2019 17:27
 Operator : JC/SP
 Sample : K1963-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-HN-MW24S-20190315

Quant Time: Mar 22 09:46:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 13:15:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	252751	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	412275	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	347508	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	139645	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	174914	45.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.90%	
35) Dibromofluoromethane	5.50	113	128803	46.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.34%	
50) Toluene-d8	8.71	98	509608	48.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.40%	
62) 4-Bromofluorobenzene	11.13	95	158491	43.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.98%	
Target Compounds						
16) Acetone	2.45	43	4470	2.070	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032119\
Data File : VX008296.D
Acq On : 21 Mar 2019 17:27
Operator : JC/SP
Sample : K1963-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-HN-MW24S-20190315

Quant Time: Mar 22 09:46:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 13:15:13 2019
Response via : Initial Calibration

