

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032119\
 Data File : VX008300.D
 Acq On : 21 Mar 2019 19:07
 Operator : JC/SP
 Sample : K1963-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-EB02-20190314

Quant Time: Mar 22 09:54:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 13:15:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	238873	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	386265	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	338507	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	137323	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	158935	44.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.36%	
35) Dibromofluoromethane	5.50	113	120093	46.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.88%	
50) Toluene-d8	8.71	98	478324	48.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.58%	
62) 4-Bromofluorobenzene	11.13	95	152876	45.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.58%	
Target Compounds						
16) Acetone	2.45	43	7245	3.551	ug/l	97
52) Toluene	8.79	92	56175	7.114	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032119\
Data File : VX008300.D
Acq On : 21 Mar 2019 19:07
Operator : JC/SP
Sample : K1963-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-EB02-20190314

Quant Time: Mar 22 09:54:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 13:15:13 2019
Response via : Initial Calibration

