

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032119\
 Data File : VX008303.D
 Acq On : 21 Mar 2019 20:22
 Operator : JC/SP
 Sample : K1860-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE105D2-20190311

Manual Integrations
 APPROVED

MMDadoda
 3/27/2019 3:31:31 PM

Quant Time: Mar 26 02:09:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 13:15:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	200754	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	316809	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	304641	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	131906	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	125277	41.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.86%	
35) Dibromofluoromethane	5.50	113	95355	44.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.92%	
50) Toluene-d8	8.71	98	408644	50.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.64%	
62) 4-Bromofluorobenzene	11.13	95	141767	51.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.42%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	23752	9.21	ug/l	97
12) 1,1-Dichloroethene	2.38	96	11125	4.57	ug/l	94
24) 1,1-Dichloroethane	3.70	63	5221	0.98	ug/l	93
27) cis-1,2-Dichloroethene	4.61	96	7148	2.38	ug/l	91
30) Chloroform	5.21	83	5037m	1.01	ug/l	
38) Carbon Tetrachloride	5.79	117	5130m	1.56	ug/l	
44) Trichloroethene	7.22	130	5840310	2187.91	ug/l	91
49) 1,4-Dioxane	7.74	88	1000	15.25	ug/l #	79
55) 1,1,2-Trichloroethane	9.21	97	2375m	0.91	ug/l	
64) Tetrachloroethene	9.34	164	6804	2.48	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032119\
Data File : VX008303.D
Acq On : 21 Mar 2019 20:22
Operator : JC/SP
Sample : K1860-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE105D2-20190311

Manual Integrations
APPROVED

MMDadoda
3/27/2019 3:31:31 PM

Quant Time: Mar 26 02:09:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031919W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 13:15:13 2019
Response via : Initial Calibration

