

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
 Data File : VX021302.D
 Acq On : 19 Mar 2021 15:16
 Operator : JC/MD
 Sample : M1690-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MC-WC

Manual Integrations
 APPROVED

MMDadoda
 3/22/2021 11:35:53 AM

Quant Time: Mar 19 15:56:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	78742	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	127547	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	102368	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	41071	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	60739	47.03	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	94.06%		
35) Dibromofluoromethane	5.464	113	43349	49.95	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.90%		
50) Toluene-d8	8.694	98	154264	48.99	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.98%		
62) 4-Bromofluorobenzene	11.118	95	50822	41.56	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	83.12%		
Target Compounds					Qvalue	
16) Acetone	2.423	43	13911	29.27	ug/l	89
17) Carbon Disulfide	2.552	76	2023	1.39	ug/l	97
25) 2-Butanone	4.641	43	2879	3.94	ug/l	91
39) Methylcyclohexane	7.435	83	1549	1.09	ug/l	94
65) Chlorobenzene	10.118	112	4097	1.95	ug/l	97
73) Isopropylbenzene	11.000	105	19455	6.38	ug/l	100
78) n-propylbenzene	11.341	91	43855	12.29	ug/l	96
79) 2-Chlorotoluene	11.400	91	4401	2.03	ug/l	87
80) 1,3,5-Trimethylbenzene	11.489	105	3586	1.39	ug/l	97
83) tert-Butylbenzene	11.753	119	5818	2.37	ug/l	76
84) 1,2,4-Trimethylbenzene	11.789	105	10453	3.98	ug/l	97
85) sec-Butylbenzene	11.930	105	29003	10.01	ug/l	99
88) 1,4-Dichlorobenzene	12.077	146	1779m	1.32	ug/l	
89) n-Butylbenzene	12.371	91	18379	7.60	ug/l #	76
91) 1,2-Dichlorobenzene	12.371	146	1522	1.18	ug/l	88
95) Naphthalene	13.818	128	8435	2.90	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
Data File : VX021302.D
Acq On : 19 Mar 2021 15:16
Operator : JC/MD
Sample : M1690-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MC-WC

Manual Integrations
APPROVED

MMDadoda
3/22/2021 11:35:53 AM

Quant Time: Mar 19 15:56:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

