

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
 Data File : VX021308.D
 Acq On : 19 Mar 2021 17:37
 Operator : JC/MD
 Sample : M1715-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 210317001-05-TRIP-BLANK

Quant Time: Mar 20 06:34:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	78989	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	6.829	114	130286	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	110358	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	44949	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	63434	48.96	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.92%
35) Dibromofluoromethane	5.458	113	44035	49.67	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.34%
50) Toluene-d8	8.694	98	157440	48.94	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.88%
62) 4-Bromofluorobenzene	11.118	95	53913	43.16	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.32%
Target Compounds						
16) Acetone	2.417	43	1704	3.57	ug/l	90
30) Chloroform	5.182	83	976	0.54	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
Data File : VX021308.D
Acq On : 19 Mar 2021 17:37
Operator : JC/MD
Sample : M1715-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
210317001-05-TRIP-BLANK

Quant Time: Mar 20 06:34:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

