

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
 Data File : VX021313.D
 Acq On : 19 Mar 2021 19:33
 Operator : JC/MD
 Sample : M1687-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AUD-1033-1042

Quant Time: Mar 20 06:36:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	77089	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	127886	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	110308	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	50343	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	60322	47.71	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.42%
35) Dibromofluoromethane	5.464	113	43052	49.47	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.94%
50) Toluene-d8	8.694	98	157270	49.81	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.62%
62) 4-Bromofluorobenzene	11.118	95	56204	45.84	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.68%
Target Compounds						
16) Acetone	2.422	43	24326	52.29	ug/l	88
25) 2-Butanone	4.634	43	87275	121.99	ug/l	92
52) Toluene	8.764	92	34658	15.78	ug/l	98
64) Tetrachloroethene	9.317	164	4689	5.69	ug/l	92
68) m/p-Xylenes	10.341	106	869	0.58	ug/l #	63
69) o-Xylene	10.682	106	1150	0.79	ug/l	96
84) 1,2,4-Trimethylbenzene	11.788	105	824	0.26	ug/l	97
95) Naphthalene	13.818	128	3378	0.95	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
Data File : VX021313.D
Acq On : 19 Mar 2021 19:33
Operator : JC/MD
Sample : M1687-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AUD-1033-1042

Quant Time: Mar 20 06:36:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

