

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
 Data File : VX021314.D
 Acq On : 19 Mar 2021 19:57
 Operator : JC/MD
 Sample : M1736-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE135D3-20210317

Quant Time: Mar 20 06:36:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	78117	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	128888	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	110758	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	46662	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	63053	49.21	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.42%
35) Dibromofluoromethane	5.464	113	44423	50.65	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.30%
50) Toluene-d8	8.694	98	158775	49.90	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.80%
62) 4-Bromofluorobenzene	11.118	95	54422	44.04	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.08%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.364	101	8263	10.12	ug/l	94
16) Acetone	2.423	43	1735	3.68	ug/l	92
27) cis-1,2-Dichloroethene	4.564	96	3573	3.50	ug/l	98
44) Trichloroethene	7.188	130	13724	14.44	ug/l	91
45) 1,2-Dichloropropane	7.488	63	4148	4.31	ug/l	92
64) Tetrachloroethene	9.318	164	16079	19.44	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
Data File : VX021314.D
Acq On : 19 Mar 2021 19:57
Operator : JC/MD
Sample : M1736-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE135D3-20210317

Quant Time: Mar 20 06:36:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

