

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
 Data File : VX021315.D
 Acq On : 19 Mar 2021 20:20
 Operator : JC/MD
 Sample : M1736-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D1-20210317

Quant Time: Mar 20 06:37:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	79629	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	130599	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	110572	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	46024	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	62525	47.87	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.74%
35) Dibromofluoromethane	5.464	113	44770	50.38	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.76%
50) Toluene-d8	8.694	98	159065	49.33	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.66%
62) 4-Bromofluorobenzene	11.118	95	55663	44.45	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.90%
Target Compounds						Qvalue
7) Trichlorofluoromethane	1.917	101	474	0.32	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.364	101	12398	14.90	ug/l	95
12) 1,1-Dichloroethene	2.358	96	6363	7.84	ug/l	85
16) Acetone	2.422	43	1788	3.72	ug/l	90
24) 1,1-Dichloroethane	3.670	63	2068	1.19	ug/l #	92
27) cis-1,2-Dichloroethene	4.564	96	3132	3.01	ug/l	99
30) Chloroform	5.170	83	897	0.50	ug/l	98
32) 1,1,1-Trichloroethane	5.464	97	676	0.42	ug/l #	38
44) Trichloroethene	7.194	130	789749	819.88	ug/l	98
55) 1,1,2-Trichloroethane	9.194	97	649	0.71	ug/l	86
64) Tetrachloroethene	9.317	164	4325	5.24	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
Data File : VX021315.D
Acq On : 19 Mar 2021 20:20
Operator : JC/MD
Sample : M1736-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D1-20210317

Quant Time: Mar 20 06:37:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

