

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
 Data File : VX021316.D
 Acq On : 19 Mar 2021 20:44
 Operator : JC/MD
 Sample : M1736-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D3-20210317

Quant Time: Mar 20 06:37:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	78477	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.829	114	127031	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	108411	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	42924	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	62423	48.50	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.00%
35) Dibromofluoromethane	5.464	113	43675	50.53	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.06%
50) Toluene-d8	8.694	98	153785	49.03	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.06%
62) 4-Bromofluorobenzene	11.117	95	53025	43.53	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.06%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.363	101	756	0.92	ug/l	Qvalue 87
16) Acetone	2.416	43	1954	4.13	ug/l	# 75
44) Trichloroethene	7.188	130	28477	30.39	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
Data File : VX021316.D
Acq On : 19 Mar 2021 20:44
Operator : JC/MD
Sample : M1736-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D3-20210317

Quant Time: Mar 20 06:37:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

