

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
 Data File : VX021317.D
 Acq On : 19 Mar 2021 21:07
 Operator : JC/MD
 Sample : M1736-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D2-20210317

Quant Time: Mar 20 06:37:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	82828	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	135383	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	115460	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	47082	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	64380	47.39	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.78%
35) Dibromofluoromethane	5.458	113	46248	50.20	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.40%
50) Toluene-d8	8.694	98	166643	49.86	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.72%
62) 4-Bromofluorobenzene	11.118	95	57661	44.42	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.84%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.364	101	21061	24.33	ug/l	96
12) 1,1-Dichloroethene	2.358	96	2235	2.65	ug/l #	83
27) cis-1,2-Dichloroethene	4.558	96	3548	3.28	ug/l	99
30) Chloroform	5.176	83	1299	0.69	ug/l #	75
38) Carbon Tetrachloride	5.758	117	1019	0.72	ug/l	91
44) Trichloroethene	7.188	130	606239	607.13	ug/l	91
49) 1,4-Dioxane	7.711	88	535	11.79	ug/l #	69
64) Tetrachloroethene	9.318	164	6193	7.18	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
Data File : VX021317.D
Acq On : 19 Mar 2021 21:07
Operator : JC/MD
Sample : M1736-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D2-20210317

Quant Time: Mar 20 06:37:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

