

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
 Data File : VX021322.D
 Acq On : 20 Mar 2021 02:38
 Operator : JC/MD
 Sample : VX0321WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0321WBL02

Quant Time: Mar 20 06:39:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	86188	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	6.823	114	142547	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	127408	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	49687	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.023	65	67464	47.72	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.44%
35) Dibromofluoromethane	5.458	113	48573	50.08	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.16%
50) Toluene-d8	8.694	98	174361	49.54	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.08%
62) 4-Bromofluorobenzene	11.118	95	60279	44.10	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.20%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
Data File : VX021322.D
Acq On : 20 Mar 2021 02:38
Operator : JC/MD
Sample : VX0321WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0321WBL02

Quant Time: Mar 20 06:39:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

