

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
 Data File : VX021326.D
 Acq On : 20 Mar 2021 05:06
 Operator : JC/MD
 Sample : M1735-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D1-20210315

Quant Time: Mar 20 07:55:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	78172	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	130879	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	112386	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	45680	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	62425	48.69	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.38%
35) Dibromofluoromethane	5.464	113	44586	50.06	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.12%
50) Toluene-d8	8.694	98	158072	48.92	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.84%
62) 4-Bromofluorobenzene	11.118	95	55991	44.62	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.24%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.182	130	5138	5.32	ug/l	95
64) Tetrachloroethene	9.318	164	600	0.71	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
Data File : VX021326.D
Acq On : 20 Mar 2021 05:06
Operator : JC/MD
Sample : M1735-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20210315

Quant Time: Mar 20 07:55:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

