

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
 Data File : VX021330.D
 Acq On : 20 Mar 2021 06:39
 Operator : JC/MD
 Sample : M1735-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D5-20210315

Quant Time: Mar 20 07:56:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	77859	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	132631	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	112619	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	46045	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	63897	50.04	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.08%
35) Dibromofluoromethane	5.464	113	45097	49.97	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.94%
50) Toluene-d8	8.694	98	161199	49.23	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.46%
62) 4-Bromofluorobenzene	11.118	95	55707	43.81	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.62%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.187	85	1184	1.27	ug/l	87
9) 1,1,2-Trichlorotrifluo...	2.364	101	9359	11.50	ug/l	96
12) 1,1-Dichloroethene	2.358	96	1794	2.26	ug/l	84
24) 1,1-Dichloroethane	3.670	63	1267	0.75	ug/l #	90
27) cis-1,2-Dichloroethene	4.564	96	1713	1.68	ug/l	97
30) Chloroform	5.176	83	1793	1.01	ug/l	96
32) 1,1,1-Trichloroethane	5.464	97	569	0.36	ug/l #	50
38) Carbon Tetrachloride	5.752	117	984	0.71	ug/l	89
44) Trichloroethene	7.188	130	110101	112.55	ug/l	93
49) 1,4-Dioxane	7.723	88	538	12.40	ug/l #	69
55) 1,1,2-Trichloroethane	9.194	97	468	0.50	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
Data File : VX021330.D
Acq On : 20 Mar 2021 06:39
Operator : JC/MD
Sample : M1735-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D5-20210315

Quant Time: Mar 20 07:56:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

