

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
 Data File : VX021332.D
 Acq On : 20 Mar 2021 07:26
 Operator : JC/MD
 Sample : M1735-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D2-20210316

Quant Time: Mar 20 07:57:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	71529	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.823	114	117296	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	101149	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	42497	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	57357	48.89	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.78%
35) Dibromofluoromethane	5.464	113	41421	51.90	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.80%
50) Toluene-d8	8.694	98	144314	49.83	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.66%
62) 4-Bromofluorobenzene	11.118	95	49917	44.38	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.76%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.187	85	1496	1.75	ug/l	94
9) 1,1,2-Trichlorotrifluo...	2.370	101	3080	4.12	ug/l	95
44) Trichloroethene	7.188	130	7263	8.40	ug/l	87
49) 1,4-Dioxane	7.717	88	517	14.45	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
Data File : VX021332.D
Acq On : 20 Mar 2021 07:26
Operator : JC/MD
Sample : M1735-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D2-20210316

Quant Time: Mar 20 07:57:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

