

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
 Data File : VX021333.D
 Acq On : 20 Mar 2021 07:50
 Operator : JC/MD
 Sample : M1735-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D3-20210316

Quant Time: Mar 22 00:51:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	75789	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	124603	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	105802	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	40001	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	60752	48.87	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.74%
35) Dibromofluoromethane	5.464	113	43077	50.81	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.62%
50) Toluene-d8	8.694	98	151823	49.35	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.70%
62) 4-Bromofluorobenzene	11.118	95	48797	40.84	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	81.68%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.187	85	595	0.66	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.364	101	3769	4.76	ug/l	96
12) 1,1-Dichloroethene	2.352	96	465	0.60	ug/l	93
27) cis-1,2-Dichloroethene	4.564	96	1213	1.22	ug/l	95
44) Trichloroethene	7.188	130	91986	100.09	ug/l	88
64) Tetrachloroethene	9.318	164	2094	2.65	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
Data File : VX021333.D
Acq On : 20 Mar 2021 07:50
Operator : JC/MD
Sample : M1735-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D3-20210316

Quant Time: Mar 22 00:51:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

