

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
 Data File : VX021334.D
 Acq On : 20 Mar 2021 08:13
 Operator : JC/MD
 Sample : M1735-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT101D-20210316

Quant Time: Mar 22 00:52:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	75765	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	125513	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	106758	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	45089	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	60930	49.03	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.06%
35) Dibromofluoromethane	5.464	113	42501	49.76	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.52%
50) Toluene-d8	8.694	98	153575	49.56	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.12%
62) 4-Bromofluorobenzene	11.118	95	53173	44.18	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.36%

Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.187	85	1382	1.53	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.364	101	13868	17.51	ug/l	95
12) 1,1-Dichloroethene	2.358	96	2499	3.24	ug/l	84
16) Acetone	2.423	43	1706	3.73	ug/l #	80
24) 1,1-Dichloroethane	3.664	63	1809	1.10	ug/l #	87
27) cis-1,2-Dichloroethene	4.564	96	4092	4.13	ug/l	96
30) Chloroform	5.176	83	1174	0.68	ug/l	92
44) Trichloroethene	7.188	130	116184	125.50	ug/l	96
49) 1,4-Dioxane	7.717	88	567	15.09	ug/l #	69
64) Tetrachloroethene	9.318	164	348	0.44	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
Data File : VX021334.D
Acq On : 20 Mar 2021 08:13
Operator : JC/MD
Sample : M1735-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT101D-20210316

Quant Time: Mar 22 00:52:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

