

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
 Data File : VX021335.D
 Acq On : 20 Mar 2021 08:36
 Operator : JC/MD
 Sample : M1735-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT101D1-20210316

Quant Time: Mar 22 00:52:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	76234	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.835	114	127487	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	108468	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	43900	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	61709	49.35	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.70%
35) Dibromofluoromethane	5.464	113	43020	49.59	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.18%
50) Toluene-d8	8.694	98	155253	49.32	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.64%
62) 4-Bromofluorobenzene	11.118	95	53284	43.59	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.18%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.187	85	860	0.94	ug/l	93
7) Trichlorofluoromethane	1.922	101	354	0.25	ug/l	93
9) 1,1,2-Trichlorotrifluo...	2.370	101	17448	21.90	ug/l	93
12) 1,1-Dichloroethene	2.358	96	6603	8.50	ug/l	89
16) Acetone	2.417	43	1716	3.73	ug/l #	75
24) 1,1-Dichloroethane	3.676	63	2291	1.38	ug/l #	86
27) cis-1,2-Dichloroethene	4.558	96	2264	2.27	ug/l	93
30) Chloroform	5.176	83	1527	0.88	ug/l	86
32) 1,1,1-Trichloroethane	5.458	97	934	0.61	ug/l #	47
38) Carbon Tetrachloride	5.752	117	2183	1.64	ug/l	94
44) Trichloroethene	7.188	130	280941	298.78	ug/l	98
55) 1,1,2-Trichloroethane	9.194	97	675	0.75	ug/l #	89
64) Tetrachloroethene	9.318	164	520	0.64	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
Data File : VX021335.D
Acq On : 20 Mar 2021 08:36
Operator : JC/MD
Sample : M1735-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT101D1-20210316

Quant Time: Mar 22 00:52:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

