

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
 Data File : VX021336.D
 Acq On : 20 Mar 2021 09:00
 Operator : JC/MD
 Sample : M1735-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D1-20210317

Quant Time: Mar 22 00:52:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	79553	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	6.829	114	130555	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	112729	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	45401	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	63251	48.48	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.96%
35) Dibromofluoromethane	5.464	113	44914	50.56	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.12%
50) Toluene-d8	8.694	98	157976	49.01	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.02%
62) 4-Bromofluorobenzene	11.118	95	55342	44.21	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.42%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.370	101	2822	3.39	ug/l	95
12) 1,1-Dichloroethene	2.358	96	696	0.86	ug/l #	93
16) Acetone	2.417	43	1730	3.60	ug/l #	87
19) Methyl tert-butyl Ether	3.176	73	2892	0.95	ug/l	91
27) cis-1,2-Dichloroethene	4.564	96	4868	4.68	ug/l	94
30) Chloroform	5.176	83	2796	1.55	ug/l	85
44) Trichloroethene	7.188	130	213634	221.86	ug/l	92
49) 1,4-Dioxane	7.723	88	717	20.79	ug/l #	79
64) Tetrachloroethene	9.318	164	11332	13.46	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
Data File : VX021336.D
Acq On : 20 Mar 2021 09:00
Operator : JC/MD
Sample : M1735-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D1-20210317

Quant Time: Mar 22 00:52:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

