

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
 Data File : VX021338.D
 Acq On : 20 Mar 2021 09:46
 Operator : JC/MD
 Sample : M1735-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D3-20210317

Quant Time: Mar 22 00:53:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	79091	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.829	114	131870	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	113350	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	44879	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	63585	49.02	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.04%
35) Dibromofluoromethane	5.464	113	45173	50.34	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.68%
50) Toluene-d8	8.694	98	159003	48.84	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.68%
62) 4-Bromofluorobenzene	11.118	95	55028	43.52	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.04%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.364	101	286456	346.56	ug/l	95
12) 1,1-Dichloroethene	2.358	96	2313	2.87	ug/l	91
16) Acetone	2.422	43	1896	3.97	ug/l	92
27) cis-1,2-Dichloroethene	4.558	96	1053	1.02	ug/l	98
44) Trichloroethene	7.188	130	23987	24.66	ug/l	92
45) 1,2-Dichloropropane	7.488	63	510	0.52	ug/l	99
64) Tetrachloroethene	9.317	164	8452	9.98	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
Data File : VX021338.D
Acq On : 20 Mar 2021 09:46
Operator : JC/MD
Sample : M1735-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D3-20210317

Quant Time: Mar 22 00:53:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

