

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
 Data File : VX021339.D
 Acq On : 20 Mar 2021 10:09
 Operator : JC/MD
 Sample : M1735-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE135D2-20210317

Quant Time: Mar 22 00:53:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	82087	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	135824	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	114674	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	48311	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	66137	49.12	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.24%
35) Dibromofluoromethane	5.464	113	46918	50.76	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.52%
50) Toluene-d8	8.694	98	163034	48.62	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.24%
62) 4-Bromofluorobenzene	11.118	95	56965	43.74	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.48%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.364	101	875	1.02	ug/l #	89
16) Acetone	2.423	43	2241	4.52	ug/l #	65
44) Trichloroethene	7.176	130	501	0.50	ug/l	96
45) 1,2-Dichloropropane	7.488	63	2277	2.25	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
Data File : VX021339.D
Acq On : 20 Mar 2021 10:09
Operator : JC/MD
Sample : M1735-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE135D2-20210317

Quant Time: Mar 22 00:53:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

