

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
 Data File : VX021343.D
 Acq On : 20 Mar 2021 11:43
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 22 00:55:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.597	0.596	0.2	93	0.00
3 P	Chloromethane	1.155	0.721	37.6#	78	0.00
4 C	Vinyl Chloride	0.608	0.612	-0.7#	93	0.00
5 T	Bromomethane	0.266	0.264	0.8	97	0.00
6 T	Chloroethane	0.372	0.374	-0.5	93	0.00
7 T	Trichlorofluoromethane	0.917	0.981	-7.0	95	0.00
8 T	Diethyl Ether	0.323	0.339	-5.0	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.523	0.531	-1.5	94	0.00
10 T	Methyl Iodide	0.456	0.459	-0.7	97	0.00
11 T	Tert butyl alcohol	0.138	0.136	1.4	95	0.00
12 CM	1,1-Dichloroethene	0.510	0.525	-2.9#	97	0.00
13 T	Acrolein	0.086	0.090	-4.7	108	0.00
14 T	Allyl chloride	1.051	1.003	4.6	88	0.00
15 T	Acrylonitrile	0.320	0.331	-3.4	95	0.00
16 T	Acetone	0.302	0.316	-4.6	101	0.00
17 T	Carbon Disulfide	1.577	1.406	10.8	92	0.00
18 T	Methyl Acetate	0.827	0.838	-1.3	99	0.00
19 T	Methyl tert-butyl Ether	1.913	2.012	-5.2	95	0.00
20 T	Methylene Chloride	0.621	0.616	0.8	96	0.00
21 T	trans-1,2-Dichloroethene	0.569	0.587	-3.2	97	0.00
22 T	Diisopropyl ether	2.018	2.144	-6.2	96	0.00
23 T	Vinyl Acetate	1.761	1.861	-5.7	93	0.00
24 P	1,1-Dichloroethane	1.090	1.130	-3.7	96	0.00
25 T	2-Butanone	0.464	0.481	-3.7	95	0.00
26 T	2,2-Dichloropropane	0.951	0.603	36.6#	57	0.00
27 T	cis-1,2-Dichloroethene	0.654	0.666	-1.8	95	0.00
28 T	Bromochloromethane	0.513	0.525	-2.3	96	0.00
29 T	Tetrahydrofuran	0.306	0.310	-1.3	94	0.00
30 C	Chloroform	1.135	1.222	-7.7#	97	0.00
31 T	Cyclohexane	0.991	0.981	1.0	92	0.00
32 T	1,1,1-Trichloroethane	1.005	1.066	-6.1	96	0.00
33 S	1,2-Dichloroethane-d4	0.820	0.835	-1.8	100	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	97	0.00
35 S	Dibromofluoromethane	0.340	0.353	-3.8	103	0.00
36 T	1,1-Dichloropropene	0.504	0.503	0.2	96	0.00
37 T	Ethyl Acetate	0.584	0.547	6.3	92	0.00
38 T	Carbon Tetrachloride	0.523	0.544	-4.0	96	0.00
39 T	Methylcyclohexane	0.559	0.544	2.7	90	0.00
40 TM	Benzene	1.388	1.438	-3.6	96	0.00
41 T	Methacrylonitrile	0.335	0.313	6.6	94	0.00
42 TM	1,2-Dichloroethane	0.599	0.616	-2.8	97	0.00
43 T	Isopropyl Acetate	0.957	0.934	2.4	92	0.00
44 TM	Trichloroethene	0.369	0.371	-0.5	95	0.00
45 C	1,2-Dichloropropane	0.373	0.388	-4.0#	96	0.00
46 T	Dibromomethane	0.273	0.275	-0.7	95	0.00
47 T	Bromodichloromethane	0.547	0.569	-4.0	95	0.00
48 T	Methyl methacrylate	0.508	0.490	3.5	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
 Data File : VX021343.D
 Acq On : 20 Mar 2021 11:43
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 22 00:55:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	95	0.00
50 S	Toluene-d8	1.234	1.237	-0.2	100	0.00
51 T	4-Methyl-2-Pentanone	0.557	0.567	-1.8	94	0.00
52 CM	Toluene	0.859	0.885	-3.0#	94	0.00
53 T	t-1,3-Dichloropropene	0.581	0.542	6.7	86	0.00
54 T	cis-1,3-Dichloropropene	0.620	0.582	6.1	85	0.00
55 T	1,1,2-Trichloroethane	0.351	0.368	-4.8	95	0.00
56 T	Ethyl methacrylate	0.565	0.569	-0.7	92	0.00
57 T	1,3-Dichloropropane	0.611	0.630	-3.1	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.309	0.303	1.9	93	0.00
59 T	2-Hexanone	0.417	0.422	-1.2	93	0.00
60 T	Dibromochloromethane	0.400	0.422	-5.5	95	0.00
61 T	1,2-Dibromoethane	0.385	0.393	-2.1	96	0.00
62 S	4-Bromofluorobenzene	0.479	0.471	1.7	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.373	0.371	0.5	95	0.00
65 PM	Chlorobenzene	1.028	1.035	-0.7	95	0.00
66 T	1,1,1,2-Tetrachloroethane	0.381	0.410	-7.6	100	0.00
67 C	Ethyl Benzene	1.901	1.918	-0.9#	93	0.00
68 T	m/p-Xylenes	0.679	0.697	-2.7	94	0.00
69 T	o-Xylene	0.660	0.674	-2.1	94	0.00
70 T	Styrene	1.107	1.144	-3.3	95	0.00
71 P	Bromoform	0.308	0.319	-3.6	93	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.714	3.796	-2.2	94	0.00
74 T	N-amyl acetate	1.884	1.818	3.5	89	0.00
75 P	1,1,2,2-Tetrachloroethane	1.233	1.269	-2.9	94	0.00
76 T	1,2,3-Trichloropropane	1.160	1.194	-2.9	103	0.00
77 T	Bromobenzene	0.881	0.892	-1.2	93	0.00
78 T	n-propylbenzene	4.343	4.404	-1.4	92	0.00
79 T	2-Chlorotoluene	2.638	2.633	0.2	92	0.00
80 T	1,3,5-Trimethylbenzene	3.143	3.238	-3.0	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.427	0.332	22.2	72	0.00
82 T	4-Chlorotoluene	3.106	3.150	-1.4	94	0.00
83 T	tert-Butylbenzene	2.992	3.079	-2.9	94	0.00
84 T	1,2,4-Trimethylbenzene	3.196	3.293	-3.0	94	0.00
85 T	sec-Butylbenzene	3.526	3.598	-2.0	93	0.00
86 T	p-Isopropyltoluene	3.231	3.365	-4.1	93	0.00
87 T	1,3-Dichlorobenzene	1.607	1.602	0.3	94	0.00
88 T	1,4-Dichlorobenzene	1.644	1.660	-1.0	94	0.00
89 T	n-Butylbenzene	2.945	2.930	0.5	90	0.00
90 T	Hexachloroethane	0.492	0.586	-19.1	106	0.00
91 T	1,2-Dichlorobenzene	1.568	1.608	-2.6	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.316	0.305	3.5	91	0.00
93 T	1,2,4-Trichlorobenzene	1.030	1.032	-0.2	93	0.00
94 T	Hexachlorobutadiene	0.432	0.406	6.0	85	0.00
95 T	Naphthalene	3.547	3.552	-0.1	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032121\
Data File : VX021343.D
Acq On : 20 Mar 2021 11:43
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 22 00:55:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.026	1.040	-1.4	90	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6