

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
 Data File : VX027579.D
 Acq On : 21 Mar 2022 12:22
 Operator : JC/MD
 Sample : VX0321MBL01
 Misc : 5.00g/10.0mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0321MBL01

Quant Time: Mar 22 05:44:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	353632	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	576692	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	516050	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	231722	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	186161	42.733	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	85.460%
35) Dibromofluoromethane	5.385	113	184786	49.803	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.600%
50) Toluene-d8	8.647	98	658255	46.883	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	93.760%
62) 4-Bromofluorobenzene	11.079	95	231871	44.931	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
Data File : VX027579.D
Acq On : 21 Mar 2022 12:22
Operator : JC/MD
Sample : VX0321MBL01
Misc : 5.00g/10.0mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0321MBL01

Quant Time: Mar 22 05:44:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

