

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
 Data File : VX027580.D
 Acq On : 21 Mar 2022 12:45
 Operator : JC/MD
 Sample : VX0321WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0321WBL01

Quant Time: Mar 22 05:44:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	380578	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	615518	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	556338	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	249119	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	198490	42.337	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	84.680%
35) Dibromofluoromethane	5.391	113	198852	50.214	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.420%
50) Toluene-d8	8.653	98	713159	47.590	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.180%
62) 4-Bromofluorobenzene	11.079	95	248138	45.050	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.100%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
Data File : VX027580.D
Acq On : 21 Mar 2022 12:45
Operator : JC/MD
Sample : VX0321WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0321WBL01

Quant Time: Mar 22 05:44:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

