

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
 Data File : VX027583.D
 Acq On : 21 Mar 2022 14:04
 Operator : JC/MD
 Sample : N1998-12DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-138DL

Quant Time: Mar 22 05:45:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	344386	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	567350	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	502603	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	224985	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	199321	46.982	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	93.960%		
35) Dibromofluoromethane	5.391	113	185635	50.856	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	101.720%		
50) Toluene-d8	8.653	98	655062	47.424	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	94.840%		
62) 4-Bromofluorobenzene	11.085	95	226286	44.571	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	89.140%		
Target Compounds						
16) Acetone	2.380	43	2293	1.326	ug/l #	81
19) Methyl tert-butyl Ether	3.117	73	371342	34.417	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
Data File : VX027583.D
Acq On : 21 Mar 2022 14:04
Operator : JC/MD
Sample : N1998-12DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-138DL

Quant Time: Mar 22 05:45:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

