

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
 Data File : VX027587.D
 Acq On : 21 Mar 2022 15:38
 Operator : JC/MD
 Sample : N1998-10DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-136DL

Quant Time: Mar 22 05:47:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	327599	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	537854	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	481012	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	216095	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	191153	47.366	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.740%		
35) Dibromofluoromethane	5.391	113	176858	51.108	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	102.220%		
50) Toluene-d8	8.653	98	622421	47.532	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	95.060%		
62) 4-Bromofluorobenzene	11.085	95	215697	44.815	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	89.640%		
Target Compounds						
11) Tert butyl alcohol	2.965	59	35862	46.627	ug/l	100
19) Methyl tert-butyl Ether	3.117	73	271915	26.493	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
Data File : VX027587.D
Acq On : 21 Mar 2022 15:38
Operator : JC/MD
Sample : N1998-10DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-136DL

Quant Time: Mar 22 05:47:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

