

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
 Data File : VX027588.D
 Acq On : 21 Mar 2022 16:01
 Operator : JC/MD
 Sample : N1998-04DL 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-130DL

Quant Time: Mar 22 05:47:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	329712	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	542523	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	482704	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	216842	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	193731	47.697	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.400%
35) Dibromofluoromethane	5.391	113	178039	51.007	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.020%
50) Toluene-d8	8.653	98	629619	47.668	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.340%
62) 4-Bromofluorobenzene	11.085	95	218623	45.032	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.060%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.965	59	51685	66.769	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	201076	19.466	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
Data File : VX027588.D
Acq On : 21 Mar 2022 16:01
Operator : JC/MD
Sample : N1998-04DL 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-130DL

Quant Time: Mar 22 05:47:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

