

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
 Data File : VX027589.D
 Acq On : 21 Mar 2022 16:25
 Operator : JC/MD
 Sample : N1998-07DL 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-133DL

Quant Time: Mar 22 05:47:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	348231	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	556594	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	513290	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	233411	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	188099	43.847	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	87.700%
35) Dibromofluoromethane	5.391	113	182700	51.019	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.040%
50) Toluene-d8	8.653	98	657428	48.515	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.040%
62) 4-Bromofluorobenzene	11.085	95	231715	46.522	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.040%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.965	59	54889	67.137	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	152992	14.023	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
Data File : VX027589.D
Acq On : 21 Mar 2022 16:25
Operator : JC/MD
Sample : N1998-07DL 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-133DL

Quant Time: Mar 22 05:47:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

