

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
 Data File : VX027590.D
 Acq On : 21 Mar 2022 16:49
 Operator : JC/MD
 Sample : N1998-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-137

Quant Time: Mar 22 05:47:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	340017	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	548585	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	483531	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	215641	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	173667	41.461	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	82.920%
35) Dibromofluoromethane	5.391	113	174378	49.406	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.820%
50) Toluene-d8	8.653	98	625140	46.806	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	93.620%
62) 4-Bromofluorobenzene	11.085	95	211429	43.069	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.140%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	6336	3.711	ug/l	96
19) Methyl tert-butyl Ether	3.117	73	335224	31.469	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
Data File : VX027590.D
Acq On : 21 Mar 2022 16:49
Operator : JC/MD
Sample : N1998-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-137

Quant Time: Mar 22 05:47:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

