

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
 Data File : VX027594.D
 Acq On : 21 Mar 2022 18:23
 Operator : JC/MD
 Sample : N2000-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-134D2-20220315

Quant Time: Mar 22 05:49:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	340181	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	561925	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	512789	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	236745	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	198152	47.28	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.56%
35) Dibromofluoromethane	5.391	113	183884	50.86	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.72%
50) Toluene-d8	8.653	98	650427	47.54	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.08%
62) 4-Bromofluorobenzene	11.085	95	229404	45.62	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.24%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	50921	15.47	ug/l	94
12) 1,1-Dichloroethene	2.313	96	1369	0.43	ug/l #	72
27) cis-1,2-Dichloroethene	4.489	96	6392	1.64	ug/l	87
44) Trichloroethene	7.129	130	409478	95.17	ug/l	100
49) 1,4-Dioxane	7.665	88	2293	22.77	ug/l	99
64) Tetrachloroethene	9.275	164	46930	11.77	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
Data File : VX027594.D
Acq On : 21 Mar 2022 18:23
Operator : JC/MD
Sample : N2000-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-134D2-20220315

Quant Time: Mar 22 05:49:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

