

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
 Data File : VX027596.D
 Acq On : 21 Mar 2022 19:11
 Operator : JC/MD
 Sample : N2000-02DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-132D4-20220315DL

Quant Time: Mar 22 05:49:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	345918	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	562338	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	513163	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	235972	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	200052	46.946	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.900%
35) Dibromofluoromethane	5.391	113	187180	51.736	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.480%
50) Toluene-d8	8.653	98	648450	47.364	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.720%
62) 4-Bromofluorobenzene	11.085	95	225228	44.758	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.520%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	2471	0.738	ug/l	97
16) Acetone	2.386	43	4248	2.446	ug/l	96
27) cis-1,2-Dichloroethene	4.495	96	1178	0.296	ug/l	84
44) Trichloroethene	7.129	130	121761	28.277	ug/l	99
64) Tetrachloroethene	9.275	164	2789	0.699	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
Data File : VX027596.D
Acq On : 21 Mar 2022 19:11
Operator : JC/MD
Sample : N2000-02DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-132D4-20220315DL

Quant Time: Mar 22 05:49:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

