

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
 Data File : VX027597.D
 Acq On : 21 Mar 2022 19:35
 Operator : JC/MD
 Sample : N2000-06DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-132D5-20220315DL

Quant Time: Mar 22 05:49:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	337035	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	544806	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	506197	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	221592	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	182866	44.04	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	88.08%
35) Dibromofluoromethane	5.391	113	183089	52.23	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.46%
50) Toluene-d8	8.653	98	630412	47.53	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.06%
62) 4-Bromofluorobenzene	11.085	95	226456	46.45	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.90%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	781	0.24	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.325	101	8439	2.59	ug/l	95
12) 1,1-Dichloroethene	2.318	96	1304	0.42	ug/l #	71
27) cis-1,2-Dichloroethene	4.501	96	1741	0.45	ug/l #	17
44) Trichloroethene	7.129	130	163606	39.22	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
Data File : VX027597.D
Acq On : 21 Mar 2022 19:35
Operator : JC/MD
Sample : N2000-06DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-132D5-20220315DL

Quant Time: Mar 22 05:49:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

