

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
 Data File : VX027598.D
 Acq On : 21 Mar 2022 19:58
 Operator : JC/MD
 Sample : N2000-13DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-139D1-20220316DL

Quant Time: Mar 22 05:50:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	335598	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	553036	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	493948	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	214016	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	191792	46.39	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.78%
35) Dibromofluoromethane	5.391	113	181459	51.00	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.00%
50) Toluene-d8	8.653	98	643649	47.80	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.60%
62) 4-Bromofluorobenzene	11.085	95	217479	43.94	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.88%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.325	101	3447	1.06	ug/l	91
12) 1,1-Dichloroethene	2.312	96	1907	0.61	ug/l #	65
27) cis-1,2-Dichloroethene	4.495	96	1430	0.37	ug/l	90
38) Carbon Tetrachloride	5.672	117	1539	0.28	ug/l	93
44) Trichloroethene	7.129	130	163726	38.66	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
Data File : VX027598.D
Acq On : 21 Mar 2022 19:58
Operator : JC/MD
Sample : N2000-13DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-139D1-20220316DL

Quant Time: Mar 22 05:50:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

