

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
 Data File : VX027600.D
 Acq On : 21 Mar 2022 20:45
 Operator : JC/MD
 Sample : N2000-10DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-134D4-20220315DL

Quant Time: Mar 22 05:50:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	332034	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	548885	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	491283	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	215511	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	190902	46.67	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.34%
35) Dibromofluoromethane	5.391	113	180622	51.15	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.30%
50) Toluene-d8	8.653	98	642426	48.07	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.14%
62) 4-Bromofluorobenzene	11.085	95	216979	44.18	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.36%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.331	101	129283	40.23	ug/l	92
44) Trichloroethene	7.129	130	19115	4.55	ug/l	91
64) Tetrachloroethene	9.275	164	8614	2.26	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
Data File : VX027600.D
Acq On : 21 Mar 2022 20:45
Operator : JC/MD
Sample : N2000-10DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-134D4-20220315DL

Quant Time: Mar 22 05:50:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

