

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
 Data File : VX027603.D
 Acq On : 21 Mar 2022 21:56
 Operator : JC/MD
 Sample : N2000-05DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-132D6-20220315DL

Quant Time: Mar 22 05:51:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	317381	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	506580	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	451095	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	209789	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	161990	41.432	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	82.860%		
35) Dibromofluoromethane	5.391	113	171335	52.569	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	105.140%		
50) Toluene-d8	8.653	98	575952	46.699	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	93.400%		
62) 4-Bromofluorobenzene	11.085	95	202465	44.663	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	89.320%		
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.325	101	4499	1.465	ug/l	90
12) 1,1-Dichloroethene	2.319	96	1238	0.420	ug/l	94
44) Trichloroethene	7.129	130	303791	78.317	ug/l	98
64) Tetrachloroethene	9.275	164	915	0.261	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032122\
Data File : VX027603.D
Acq On : 21 Mar 2022 21:56
Operator : JC/MD
Sample : N2000-05DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-132D6-20220315DL

Quant Time: Mar 22 05:51:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

