

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032123\
 Data File : VX034655.D
 Acq On : 21 Mar 2023 15:27
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: Mar 22 02:05:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	250844	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	419613	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	360214	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	154751	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	165629	45.387	ug/l	0.01
Spiked Amount 50.000	Range 78 - 117		Recovery =	90.780%		
35) Dibromofluoromethane	5.385	113	131225	47.365	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.740%		
50) Toluene-d8	8.647	98	502654	48.601	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.200%		
62) 4-Bromofluorobenzene	11.079	95	172256	41.234	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	82.460%#		
Target Compounds						
11) Tert butyl alcohol	2.940	59	4332	5.700	ug/l #	82
17) Carbon Disulfide	2.514	76	2408	0.312	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032123\
Data File : VX034655.D
Acq On : 21 Mar 2023 15:27
Operator : JC/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK

Quant Time: Mar 22 02:05:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

