

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032123\
 Data File : VX034656.D
 Acq On : 21 Mar 2023 15:50
 Operator : JC/MD
 Sample : 01946-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NWB-1896

Quant Time: Mar 22 02:06:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	249077	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	419692	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	366889	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	179191	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	171459	47.318	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.640%
35) Dibromofluoromethane	5.391	113	108313	39.088	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	78.180%
50) Toluene-d8	8.647	98	496191	47.967	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.940%
62) 4-Bromofluorobenzene	11.085	95	198423	47.489	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.980%
Target Compounds						
					Qvalue	
16) Acetone	2.441	43	77925	45.653	ug/l	93
18) Methyl Acetate	2.733	43	388109	69.588	ug/l	100
25) 2-Butanone	4.629	43	12361	5.061	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032123\
Data File : VX034656.D
Acq On : 21 Mar 2023 15:50
Operator : JC/MD
Sample : 01946-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NWB-1896

Quant Time: Mar 22 02:06:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

