

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032124\
 Data File : VX040750.D
 Acq On : 21 Mar 2024 18:09
 Operator : JC/MD
 Sample : PB159677ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB159677ZHE#01

Quant Time: Mar 22 02:58:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:41:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	44031	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	77832	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	72822	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	39105	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	37190	50.232	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 100.460%			
35) Dibromofluoromethane	5.385	113	26632	45.729	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 91.460%			
50) Toluene-d8	8.646	98	87580	43.919	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 87.840%#			
62) 4-Bromofluorobenzene	11.079	95	38930	49.413	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 98.820%			
Target Compounds						Qvalue
11) Tert butyl alcohol	2.946	59	2539	21.416	ug/l	99
16) Acetone	2.373	43	9387	35.796	ug/l	97
20) Methylene Chloride	2.800	84	3534	6.165	ug/l	87
25) 2-Butanone	4.562	43	9560	23.275	ug/l	96
43) Isopropyl Acetate	6.336	43	38208	24.988	ug/l	97
59) 2-Hexanone	9.433	43	4470	6.085	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032124\
Data File : VX040750.D
Acq On : 21 Mar 2024 18:09
Operator : JC/MD
Sample : PB159677ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB159677ZHE#01

Quant Time: Mar 22 02:58:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:41:46 2024
Response via : Initial Calibration

