

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032124\
 Data File : VX040754.D
 Acq On : 21 Mar 2024 19:40
 Operator : JC/MD
 Sample : PB159677ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB159677ZHE#05

Quant Time: Mar 22 03:00:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:41:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	41063	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.763	114	74356	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	66690	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	36916	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	34034	49.292	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.580%
35) Dibromofluoromethane	5.385	113	24791	44.558	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.120%
50) Toluene-d8	8.647	98	84949	44.591	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.180%#
62) 4-Bromofluorobenzene	11.079	95	34938	46.419	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.840%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.947	59	2297	20.775	ug/l	98
16) Acetone	2.380	43	9393	38.407	ug/l	97
20) Methylene Chloride	2.794	84	3472	6.495	ug/l	94
25) 2-Butanone	4.562	43	10045	26.223	ug/l	94
43) Isopropyl Acetate	6.336	43	36131	24.735	ug/l	98
59) 2-Hexanone	9.433	43	4757	6.779	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032124\
Data File : VX040754.D
Acq On : 21 Mar 2024 19:40
Operator : JC/MD
Sample : PB159677ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB159677ZHE#05

Quant Time: Mar 22 03:00:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:41:46 2024
Response via : Initial Calibration

