

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032124\
 Data File : VX040759.D
 Acq On : 21 Mar 2024 21:35
 Operator : JC/MD
 Sample : PB159677ZHE#10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB159677ZHE#10

Quant Time: Mar 22 03:02:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:41:46 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	41663	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	74160	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	72324	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	36498	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	35095	50.097	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 100.200%		
35) Dibromofluoromethane	5.385	113	25188	45.391	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 90.780%		
50) Toluene-d8	8.647	98	88444	46.548	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 93.100%		
62) 4-Bromofluorobenzene	11.079	95	35934	47.868	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 95.740%		
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	26581	107.123	ug/l	97
20) Methylene Chloride	2.794	84	2360	4.351	ug/l	97
25) 2-Butanone	4.556	43	2992	7.698	ug/l	99
43) Isopropyl Acetate	6.336	43	39249	26.940	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032124\
Data File : VX040759.D
Acq On : 21 Mar 2024 21:35
Operator : JC/MD
Sample : PB159677ZHE#10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB159677ZHE#10

Quant Time: Mar 22 03:02:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:41:46 2024
Response via : Initial Calibration

